



TEST REPORT TUCHENG

Date: Oct.07,2020 Date of Receipt: Sep.22,2020

Report No.: TF909434 Quantity: 1PC Page Order/Pages: (P1/2) Ref. No.: NIL

Report Title: Item: Isolation Gowns


Address:

Test Items		Test Results				Test Methods
		Main Body	Seam At			
			Sleeve	Armhole	Shoulder	
Water Resistance	1	67.8	59.1	31.8	58.5	ANSI/AAMI PB70:2012
Hydrostatic Pressure Test (cmH2O)	2	69.0	41.4	44.4	59.8	AATCC 127-2018
	3	77.2	55.2	31.4	46.0	The rate of water
	Ave	71.3	51.9	35.9	54.8	pressure:60± 3cmH2O/min
Water Resistance	1	<0.1	<0.1	<0.1	<0.1	ANSI/AAMI PB70:2012
Impact Penetration Test (g)	2	<0.1	<0.1	<0.1	<0.1	AATCC 42-2017 Type II
	3	<0.1	<0.1	<0.1	<0.1	Environmental Condition:
	Ave	<0.1	<0.1	<0.1	<0.1	27°C ,60%RH

Note: Sample description is given by the client: ISOGOWNCIN3-11.5-11-AX Isolation Gown

Note: As the remaining sample was asked to return along with the test report, re-testing would not be possible.

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.  
2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.  
3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of  
Taiwan Textile Research Institute

*Jui-hung kao*

Director,  
Department of Testing and  
Certification

Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Oct. 07, 2020 Date of Receipt: Sep. 22, 2020

Report No.: TF909434 Quantity: 1PC Page Order/Pages: (P2/2) Ref. No.: NIL

Report Title: Item: Isolation Gowns

Address:



- Note: 1. This report is only responsible for the submitted sample(s), which will be kept for one month period.
2. This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
3. The test report should not be used for public advertisement and commercial promotion.

Authorized by president of  
Taiwan Textile Research Institute

*Jui-hung kao*

Director,  
Department of Testing and  
Certification

Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108 , +886-2-22689839